

ComfortCushion is a firm cushion in anti-slip material that is placed under the patient's knees to prevent the patient from sliding forward when sitting up in bed. While the patient is prevented from sliding forward, the cushion also prevents stretching at the back of the thighs, which can otherwise occur when the legs are straight. ComfortCushion is made of a material with closed cells, which makes it light and easy to keep clean.

SystemRoMedic is a series of different models of transfer assistive devices. The choice of assistive devices depends on the transfer situation and the patient's functional capability and needs.



Functional inspection

Visual inspection

Check the condition and function of the assistive device regularly. Always inspect the product after laundering. Check to ensure that seams and material are free from damage. Check to ensure that fabric is not worn or faded. If there are signs of wear, the product must be discarded.





Always read the manual

Always read the manuals for all assistive devices used during a transfer.

Keep the manual where it is accessible to users of the product.

Do not leave the patient unattended during a transfer situation.

Placement of the ComfortCushion:



Placed under the patient's knees, when the patient is lying down, to prevent the patient from sliding downward when the head end of the bed is raised.

When raising the head end of the bed, ensure that pressure against the shoulders is alleviated, either by leaning the patient forward or by placing a friction-reducing device behind the shoulders.

NOTE! There may be a risk for contractures. Do not leave the cushion under the knees for too long.

Material:

Polyethylen

Care of the product:

Clean with dishwashing liquid or disinfectant. Wipe with a cloth.

Note: Do not subject the product to hard pressure or impact. This may leave indentations in the material.



P O Box 4791 Ferndown Dorset BH22 9WB
Tel: 07857 603031 Fax: 01202 877240
Email: info@amilakesouthern.co.uk